## Semi-continuous Benzene Monitoring in Burlington, Vermont July 2007-June 2008



Robert C. Lacaillade, Vermont APCD NESCAUM Monitoring Assessment Committee Putney, VT 11/6/2008

## Syntec Spectras GC955 Series 600 BTEX Analyzer


- Manufacturer: *Synspec* (The Netherlands)
  - U.S. Distributor/Customer Support: Wilbur Technical Services, LLC
- Computer controlled-embedded PC running proprietary software
  - Windows XP, 40 Gigabyte HD, PC-Anywhere
  - 19" Standard Rack Mount
  - 10" integrated monitor, External keyboard & mouse
  - Various communication/data options (USB, modem, ethernet etc..)



### **Design Specifications**



**GC955 Inside Top view** 



- Carrier gas- N<sub>2</sub> (Ultra High Purity)
- Preconcentrator- Tenax GR
- GC-Based
  - Column: 15 m x .32mm (ID)
    - 95% dimethylpolysiloxane;
      5% diphenylpolysiloxane
    - 2-meter "stripper" column
    - 13-meter "analysis" column
- PID- 10.6 eV, 50 µl measurement cell
- "Semi-continuous"; 15-minute analysis run time per sample
- 2 operating modes:
  - sample Injection
  - sampling/Analysis
- Design keys
  - 10-port valve
  - preconcentrator
  - sampling piston

## Installation


- Burlington Trailer
  - In City Parking lot on corner of 2 busy streets
  - Near 3 gas stations, fire station, news paper publisher
- Connected to common glass sample manifold
  - 1/8" Silco-coated SS tubing
  - 2µ sintered-SS particulate inlet filter
- Internet connection via dedicated DSL line
  - Allows offsite communication, data review/download
  - Remote control via PC-anywhere
- LAN connection with trailer PC

## Calibration

- Calibration gas:
  - Spectra or Scott Cylinder
    - BTEX compounds at 1 ppmv nominal, Balance Nitrogen
- ZAS: TEI Model 111
- Dilution Calibrator:
  - Environics Model 6103
    - 0-10 sccm gas MFC
    - 0-20 lpm Air MFC



## **Benzene Calibration**




- 4-point calibration, 0-5 ppbv range
  - ppbv input concentration levels 0.25, 0.75, 2.00, 4.75
- linear regression (not thru 0, auto-linearization not activated)
  - $R^2 = 0.9998$
  - y-intercept becomes lowest reportable level (0.06 ppbv)


# Data Handling

- Chromatogram results from each 15-min sample run saved to unique data file
  - Saved files can be loaded and reviewed in main screen for review and final validation
- ppbv results for each 15-minute sample are written to a text file
  - Each text file contains 1 month of data
  - Text file is retrieved using PC-Anywhere
  - Text file imported to EXCEL and ACCESS for review, processing, validation etc.....
  - 15-min values used to generate 1-hour averages
    - 75% data averaging rule (min. of 3, 15-min values for 1-hr avg)

# Data Validation

- Text file of 15-min results imported to Excel/Access for review:
  - missing, suspect or elevated values, retention time verification
  - Flagging and validation
- Review chromatograms of selected runs
  - Verify retention time, peak area, identification, concentration
- Review results graphically
  - Time series, Fingerprint plots
  - Scatter Plot of benzene/toluene ratios for review of suspect values:






## **Quality Assurance**

- Weekly performance/precision audit
  - 2 ppbv input
  - DQO +/-15% (85%-115% recovery)
- Biannual accuracy audit
  - second source standard; mid-range
  - DQO +/-20%
- Data capture
  - Quarterly DQO: 75%


#### Summary of Weekly Audit of Synspec GC955 (@ 2 ppbv) July 2007- June 2008




Benzene CV= 5.3% (40CFRPart58, App.A, 7-1-08 Rev.) Eq. 2) Upper/Lower Probability Limits= 3.6%/-14.3% (Eq. 6 & 7) Mean precision error for project period: Benzene=-5.3% Toluene=-6.9%

### Second Source Benzene Performance Audit Results For Synspec GC955

Audit Date	Input (ppbv)	Response (ppbv)	% Diff.
5/4/07 (Source: Spectra)	1.4	1.1	-21
<b>8/31/07</b> (Source:ERG)	2.7	3.0	11
4/24/08 (Source:Spectra)	3.0	3.0	0
8/2/08 (Source:Spectra)	2.0	2.0	0




#### Semi-Continuous (GC/PID) Compared to Whole Air Sample (GC/MS) (24-hour Average Concentrations)



Note: Synspec 24-hr averages based on mean of 96 separate 15-minute values for each day; TO-15 results based on individual analysis of 24-hour whole-air canister sample


### Semi-Continuous (GC/PID) Compared to Whole Air Sample (GC/MS) Continued

(6-hour Average Concentrations)




*Note: Synspec 6-hr averages based on mean of 24 separate 15-minute values for each day; TO-15 results based on individual analysis of 6-hour whole-air canister sample* 

#### Semi-Continuous (GC/PID) Compared to Whole Air Sample (GC/MS) Continued




Note: Synspec 24-hr and 6-hr averages based on mean of 96 or 24 separate 15-minute values for each sample period, respectively; TO-15 results based on an individual analysis of 24-hour or 6-hour whole-air canister sample

#### Semi-Continuous (GC/PID) Compared to Whole Air Sample (GC/MS) Continued




(40CFRPart58, AppendixA, 7-1-08 Revision)


CV=13.2%




#### Semi-Continuous (GC/PID) Compared to Whole Air (GC/MS) Continued



CV=15.6% (40CFRPart58, AppendixA, Eq. 11, 7-1-08 Revision)



#### Semi-Continuous (GC/PID) Compared to Whole Air (GC/MS) Continued



Mean Difference= -0.04 ppbv Range= -0.13 to 0.04 ppbv

#### Conclusions/Observations on GC955 from Burlington Benzene Study

#### • Analyzer installation and operation straight forward

- Field ready
- Windows based software
- Calibration/audit gas available from multiple vendors
- Ability to use existing station diluter & ZAS
- Time/experience necessary for familiarity with program and establishing proper integration parameters and proper review/validation of results
  - Vendor training was helpful

#### Consumables and maintenance manageable

- Nitrogen consumption: 1 large (size 300 Airgas) cylinder every 6 weeks
- Particulate filter: quarterly
- Critical analyzer maintenance able to be performed in-house
  - Sample preconcentrator and carrier gas filters exchanged annually
  - PID cleaned biannually or as-needed

#### Performance enhancements

- Trace level blender necessary for calibration & audits
- Silco-coated ss tubing for sample inlet
- Linearity achieved without auto-linearization function enabled

#### Acceptable data quality achievable

- with regular performance checks and calibrations
- with thorough data review, management and validation
  - Although it's time intensive and can be complicated

#### Conclusions/Observations on GC955 from Burlington Benzene Study (Continued)

#### • Analyzer is well designed and reliable

- For Burlington location, measurement range appropriate for observed short-term benzene concentrations
  - Very few 15-minute measurements recorded > 5 ppbv
- After startup and imbedded PC issues, GC955 operated with minimal down time during entire project period
  - 94% Data capture for project period

#### Enhancement to Vermont Air Toxics Monitoring Program

- Provides real-time Benzene concentrations that <u>on average</u> correlate well with collocated TO-15 canister analyses but indicate a negative bias mainly +/- 0.05 ppbv (average difference for 24-hr average comparisons over the project= -0.04 ppbv; n=30)
- Benzene ppbv minimum detection level is comparable to TO-15(<0.2 ppbv)
- All QAPP DQO's were met for this project
- Overall, meets applicable EPA performance-based criteria for NATTS and TO methods
- Continue to optimize calibration, performance and data quality and evaluate relationship to TO-15 results.

#### Acknowledgement:

### Thanks to





Jenny Berschling

John Simone

\*\*For all of their hard work, diligence and invaluable efforts which significantly contributed to the success of this project!\*\*